

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	General Medical Devices Hypo-/hyperthermal device	
Name and address of the applicant	Hilotherm GmbH Wittumweg 38 88260 Argenbühl-Eisenharz GERMANY	
Name and address of the manufacturer	Hilotherm GmbH Wittumweg 38, 88260 Argenbühl-Eisenharz, GERMANY	
Name and address of the factory	Hilotherm GmbH Wittumweg 38, 88260 Argenbühl-Eisenharz, GERMANY	
Ratings and principal characteristics	Rated voltage	230 V _{AC}
	Rated frequency	50 Hz
	Power consumption	320 VA
	Protection class	I
	Degree of Protection	IP20
	Type of applied part	B
Trade mark (if any)	HILOTHERM	
Model/type Ref.	Hilotherm HT02	
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 IEC 60601-1:2005/AMD1:2012 IEC 60601-1-6:2010 IEC 60601-1-6:2010/AMD1:2013	
as shown in the Test Report Ref. No. which forms part of this certificate	713174362	

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 096215 0008 Rev. 00

Date, 2020-11-16



(Paul Bielawski)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service